

Technical Data Sheet

BUV737 Mouse Anti-Human HLA-DQ

Product Information

Material Number:	749003
Size:	50 µg
Clone:	SK10
Alternative Name:	MHC class II HLA-DQ; MHC HLA-DQ
Reactivity:	Human (Tested in Development)
Isotype:	Mouse IgG1, κ
Immunogen:	Tonsillar B lymphocytes (non-E-rosetting cells)
Application:	Flow cytometry (Qualified)
Concentration:	0.2 mg/ml
Storage Buffer:	Aqueous buffered solution containing ≤0.09% sodium azide.

Description

The SK10 monoclonal antibody specifically recognizes a class II major histocompatibility complex (MHC) antigen, with a molecular weight of 26 to 34 kDa, distinct from HLA-DR3-5 and HLA-DP5 antigens. Anti-HLA-DQ recognizes a common polymorphic epitope present on HLA-DQ molecules of cells expressing DQw1 and DQw3 (usually associated with DR1, DR2, DR4, DR5, DRw8, DRw9, and DRw10) and absent from DQw2 (usually associated with DR3, and DR7). Therefore, the SK10 antibody not react with cells from DR3 or DR7 homozygotes or DR3/7 heterozygotes. The HLA-DQ antigen is present on approximately 10% of peripheral blood lymphocytes. HLA-DQ reacts weakly with most peripheral blood monocytes and mitogen-stimulated T-lymphocyte blasts. The SK10 antibody reacts with virtually all B-cell lines, some myelomas, and some myeloid leukemias, but only rarely with T-lymphocyte tumors.

The antibody was conjugated to BD Horizon™ BUV737 which is part of the BD Horizon Brilliant™ Ultraviolet family of dyes. This dye is a tandem fluorochrome of BD Horizon BUV395 with an Ex Max of 348-nm and an acceptor dye with an Em Max at 737-nm. BD Horizon Brilliant BUV737 can be excited by the ultraviolet laser (355 nm) and detected with a 740/35 filter. Due to the excitation of the acceptor dye by other laser lines, there may be significant spillover into channels detecting Alexa Fluor® 700-like dyes (eg, 712/20-nm filter).

Due to spectral differences between labeled cells and beads, using BD™ CompBeads can result in incorrect spillover values when used with BD Horizon BUV737 reagents. Therefore, the use of BD CompBeads or BD CompBeads Plus to determine spillover values for these reagents is not recommended. Different BUV737 reagents (eg, CD4 vs. CD45) can have slightly different fluorescence spillover therefore, it may also be necessary to use clone specific compensation controls when using these reagents.

Preparation and Storage Section

Store undiluted at 4°C and protected from prolonged exposure to light. Do not freeze. The monoclonal antibody was purified from tissue culture supernatant or ascites by affinity chromatography. The antibody was conjugated with BD Horizon BUV737 under optimal conditions that minimize unconjugated dye and antibody.

Recommended Assay Procedure

For optimal and reproducible results, BD Horizon Brilliant Stain Buffer should be used anytime two or more BD Horizon Brilliant dyes (including BD OptiBuild Brilliant reagents) are used in the same experiment. Fluorescent dye interactions may cause staining artifacts which may affect data interpretation. The BD Horizon Brilliant Stain Buffer was designed to minimize these interactions. More information can be found in the Technical Data Sheet of the BD Horizon Brilliant Stain Buffer (Cat. No. 563794).

Suggested Companion Products

Catalog Number	Name	Size	Clone
612758	BUV737 Mouse IgG1, κ Isotype Control X40 RUO	50 µg	
349202	Lysing Solution 10X Concentrate IVD	100 NA	
564219	Human BD Fc Block™ RUO	50 mg	
554656	Stain Buffer (FBS) RUO	500 mL	
554657	Stain Buffer (BSA) RUO	500 mL	

Product Notices

1. This antibody was developed for use in flow cytometry.
2. The production process underwent stringent testing and validation to assure that it generates a high-quality conjugate with consistent performance and specific binding activity. However, verification testing has not been performed on all conjugate lots.
3. Researchers should determine the optimal concentration of this reagent for their individual applications.
4. An isotype control should be used at the same concentration as the antibody of interest.
5. Caution: Sodium azide yields highly toxic hydrazoic acid under acidic conditions. Dilute azide compounds in running water before discarding to avoid accumulation of potentially explosive deposits in plumbing.
6. For fluorochrome spectra and suitable instrument settings, please refer to our Multicolor Flow Cytometry web page at wwwbdbiosciences.com/colors.
7. Please refer to wwwbdbiosciences.com/us/s/resources for technical protocols.
8. BD Horizon Brilliant Stain Buffer is covered by one or more of the following US patents: 8,110,673; 8,158,444; 8,575,303; 8,354,239.
9. BD Horizon Brilliant Ultraviolet 737 is covered by one or more of the following US patents: 8,110,673; 8,158,444; 8,227,187; 8,575,303; 8,354,239.

References

- Brodsky FM. A matrix approach to human class II histocompatibility antigens: reactions of four monoclonal antibodies with the products of nine haplotypes.. *Immunogenetics*. 1984; 19(3):179-94. (Biology: Flow cytometry).
- Chen YX, Evans RL, Pollack MS, et al. Characterization and expression of the HLA-DC antigens defined by anti-Leu 10.. *Hum Immunol*. 1984; 10(4):221-35. (Clone-specific: Flow cytometry).
- Crompton MF, Bodmer JG, Bodmer WF, Heyes JM, Lindsay J, Rudd CE. Biochemistry of class II antigens: workshop report. In: :29.
- Edwards JA, Durant BM, Jones DB, Evans PR, Smith JL. Differential expression of HLA class II antigens in fetal human spleen: relationship of HLA-DP, DQ, and DR to immunoglobulin expression.. *J Immunol*. 1986; 137(2):490-7. (Biology: Flow cytometry).
- Gonwa TA, Frost JP, Karr RW. All human monocytes have the capability of expressing HLA-DQ and HLA-DP molecules upon stimulation with interferon-gamma.. *J Immunol*. 1986; 137(2):519-24. (Biology: Flow cytometry).
- Robbins PA, Evans EL, Ding AH, Warner NL, Brodsky FM. Monoclonal antibodies that distinguish between class II antigens (HLA-DP, DQ, and DR) in 14 haplotypes.. *Hum Immunol*. 1987; 18(4):301-13. (Biology: Flow cytometry).

BD Biosciences

bdbiosciences.com

United States 877.232.8995	Canada 888.268.5430	Europe 32.53.720.550	Japan 0120.8555.90	Asia Pacific 65.6861.0633	Latin America/Caribbean 0800.771.7157
-------------------------------	------------------------	-------------------------	-----------------------	------------------------------	--

For country contact information, visit bdbiosciences.com/contact

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for a patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton Dickinson and Company is strictly prohibited. For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

©2020 BD. All rights reserved. Unless otherwise noted, BD, the BD Logo and all other trademarks are the property of Becton, Dickinson and Company or its affiliates.

