




Technical Data Sheet

Phospho ZAP70 (Y319) Flex Set

Product Information

Material Number:	560011
Size:	100 Tests
Bead Position:	B8
Assay Range:	3.9-1,000 Units/mL
Reactivity:	QC Testing: Human Tested in Development: Mouse, Rat

Component Description: Phospho ZAP70 (Y319) PE Detection Reagent
 Component Mat. No: 51-9003912
 Component Storage Buffer: Aqueous buffered solution containing BSA and $\leq 0.09\%$ sodium azide.

Component Description: Phospho ZAP70 (Y319) Standard
 Component Mat. No: 51-9005187
 Component Storage Buffer: Lyophilized in an aqueous buffered solution containing BSA and ProClin™ 150.

Component Description: Phospho ZAP70 (Y319) Capture Bead B8
 Component Mat. No: 51-9005189
 Component Storage Buffer: Aqueous buffered solution containing fetal bovine serum and $\leq 0.09\%$ sodium azide.




Figure 1. Example BD CBA Phospho ZAP70 (Y319) Flex Set standard curve. Data acquired on a BD FACSAria bioanalyzer and analyzed using FCAP Array Software.

Description

The BD™ CBA Phospho ZAP70 (Y319) Flex Set is a bead-based immunoassay capable of measuring human, mouse or rat ZAP70, a protein tyrosine kinase, which has been tyrosine-phosphorylated at Tyr-319 in denatured cell lysate samples. Human reactivity was determined by testing cell lysates with the BD CBA Phospho ZAP70 (Y319) Flex Set. Reactivity with mouse and rat samples was determined by western blot for each of the antibodies used in the BD CBA Phospho ZAP70 (Y319) Flex Set. There are several publications evaluating the biology and function of ZAP70. For more information on bead-based immunoassays, refer to the product insert for the BD CBA Cell Signaling Master Buffer Kit (Cat. No. 560005 or 560006).

Preparation and Storage

This BD™ CBA Flex Set contains one vial of each component listed above. All components of this flex set have been formulated to a 50x concentration to ensure product performance when multiplexed. Store at 4°C. Protect Capture Beads and the PE Detection Reagent from prolonged exposure to light.

The Phospho ZAP70 (Y319) Standard provided in this Flex Set is lyophilized and the standard sphere should be transferred to a 1.5 mL microfuge tube for reconstitution. Reconstitute the standard with 0.1 mL Assay Diluent from the BD CBA Cell Signaling Master Buffer Kit (Cat. No. 560005/560006), warm to 37 °C and vortex prior to use. After reconstitution, the standard concentration is 50,000 Units/mL and is stable for 3 months when stored at 4 °C. When using reconstituted standard, warm to 37 °C and vortex to mix thoroughly before use.

Application Notes

Recommended Assay Procedure: The BD CBA Phospho ZAP70 (Y319) Flex Set must be used in conjunction with a BD CBA Cell Signaling Master Buffer Kit (Cat. No. 560005, 100 tests, or 560006, 500 tests) and a flow cytometer. Detailed instructions on the use of this product can be found in the manual for the BD CBA Cell Signaling Master Buffer Kit. When following the directions in the Master Buffer Kit, the top standard

BD Biosciences

bdbiosciences.com

United States	Canada	Europe	Japan	Asia Pacific	Latin America/Caribbean
877.232.8995	800.268.5430	32.2.400.98.95	0120.8555.90	65.6861.0633	55.11.5185.9995

For country contact information, visit bdbiosciences.com/contact

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

Unless otherwise noted, BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2014 BD



point for the BD CBA Phospho ZAP70 (Y319) Flex Set will be 1,000 Units/mL. An example standard curve is shown in Figure 1.

The BD CBA Phospho ZAP70 (Y319) Flex Set **cannot** be used in the same assay well with the following BD CBA Cell Signaling Flex Set reagents:

Flex Set	Bead Position	Catalog Number
BD CBA Phospho Akt2 (T309) Flex Set	A5	560146
BD CBA Phospho JNK1/2 (T202/Y204) Flex Set	B5	560013

The BD CBA Phospho ZAP70 (Y319) Flex Set should not be used in the same assay well with any non-BD CBA Cell Signaling Flex Set reagents (such as BD CBA Human or Mouse Soluble Protein Flex Sets) **nor with any BD CBA Total Protein Cell Signaling Flex Set reagents**. For an updated assay compatibility chart for the BD CBA Cell Signaling Flex Sets, please refer to the BD CBA Flex Set System homepage at <http://www.bdbiosciences.com/cbasetup>.

Performance

Limit of Detection: The theoretical limit of detection is 0.75 Units/mL and was determined by evaluating the estimated result of the average MFI of the negative control (0 Units/mL, $n=30$) + 2 standard deviations.

Specificity		Inter-Assay Reproducibility			Intra-Assay Reproducibility		
		Mean (Units/ml)	Standard Deviation	%CV	Mean (Units/ml)	Standard Deviation	%CV
Phospho ZAP70 (Y319)	Sample 1	30.5	1.5	5%	30.3	1.2	4%
	Sample 2	123.5	8.7	7%	121.1	5.0	4%
	Sample 3	500.5	34.0	7%	489.5	14.8	3%

Reproducibility: The inter-assay and intra-assay reproducibility were determined for the BD CBA Phospho ZAP70 (Y319) Flex Set by evaluating ten replicates of three different sample levels (intra-assay) and two replicates of three different sample levels from four separate experiments (inter-assay).

Lysate Dilution	Phospho ZAP70 (Y319)	
	Detected (Units/ml)	% of Expected
Neat	913.9	100%
1 : 2	497.6	109%
1 : 4	255.2	112%
1 : 8	117.7	103%

Linearity: An activated cell lysate was serially diluted to determine the linearity of the assay.

Product Notices

- Source of all serum proteins is from USDA inspected abattoirs located in the United States.
- ProClin is a trademark of Rohm and Haas Company.
- Caution: Sodium azide yields highly toxic hydrazoic acid under acidic conditions. Dilute azide compounds in running water before discarding to avoid accumulation of potentially explosive deposits in plumbing.
- Warning: CBA lyophilized standard contains 32.07% sodium dodecyl sulfate (w/w) and 0.01% (w/w) of a CMIT/MIT mixture (3:1), which is a mixture of: 5-chloro-2-methyl-4-isothiazolin-3-one [EC No 247-500-7] and 2-methyl-4-isothiazolin-3-one [EC No 220-239-6] (3:1). Hazard statement: May be harmful if swallowed. Causes skin irritation. Causes serious eye irritation. May cause an allergic skin reaction. Precautionary statements: Wear protective gloves/eye protection. Wear protective clothing. Avoid breathing mist/vapours/spray. If skin irritation or rash occurs: Get medical advice/attention. IF ON SKIN: Wash with plenty of water. Dispose of contents/container in accordance with local/regional/national/international regulations.

BD Biosciences

bdbiosciences.com

United States	Canada	Europe	Japan	Asia Pacific	Latin America/Caribbean
877.232.8995	800.268.5430	32.2.400.98.95	0120.8555.90	65.6861.0633	55.11.5185.9995

For country contact information, visit bdbiosciences.com/contact

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

Unless otherwise noted, BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2014 BD

