



# BD FACS™ Accudrop Beads

Catalog No.	Volume	Conc.	Tests	Volume/Test
345249	1.5 mL	1.0 × 10 <sup>8</sup> beads/mL	25	1 drop in 0.5 mL PBS

## RESEARCH APPLICATIONS

Research applications include:

- Setting the drop delay on BD FACSVantage™ SE flow cytometers equipped with the BD FACS™ Accudrop option, or BD FACSaria™ flow cytometers

## DESCRIPTION

This product contains a single population of 6-µm particles. Every particle contains a fluorophore that is excited at 670 nm and emits at 750 nm.

The particles are supplied in 1.5 mL of water containing 0.05% Tween® 20 and 2 mM sodium azide.

## PROCEDURE

1. Shake bead vial vigorously or vortex briefly before use.
2. Dilute 1 drop of beads in 0.5 mL phosphate-buffered saline (PBS).
3. Use as directed in the instrument or option user's guide.

**NOTE** Use diluted particles within 24 hours after dilution.

## HANDLING AND STORAGE

Store vials at 2°C–8°C. Protect from exposure to light. Each reagent is stable until the expiration date shown on the bottle label when stored as directed.

## WARRANTY

Unless otherwise indicated in any applicable BD general conditions of sale for non-US customers, the following warranty applies to the purchase of these products.

THE PRODUCTS SOLD HEREUNDER ARE WARRANTED ONLY TO CONFORM TO THE QUANTITY AND CONTENTS STATED ON THE LABEL OR IN THE PRODUCT LABELING AT THE TIME OF DELIVERY TO THE CUSTOMER. BD DISCLAIMS HEREBY ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING WARRANTIES OF MERCHANTABILITY AND FITNESS FOR ANY PARTICULAR PURPOSE AND NONINFRINGEMENT. BD'S SOLE LIABILITY IS LIMITED TO EITHER REPLACEMENT OF THE PRODUCTS OR REFUND OF THE PURCHASE PRICE. BD IS NOT LIABLE FOR PROPERTY DAMAGE OR ANY INCIDENTAL OR CONSEQUENTIAL DAMAGES, INCLUDING PERSONAL INJURY, OR ECONOMIC LOSS, CAUSED BY THE PRODUCT.

**For Research Use Only. Not for use in diagnostic or therapeutic procedures.**

Becton, Dickinson and Company  
BD Biosciences  
2350 Qume Drive  
San Jose, CA 95131 USA



**TRADEMARKS AND  
LICENSE INFORMATION**

Tween is a registered trademark of Sigma-Aldrich.

BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2015 BD

This product is provided under a license agreement between Life Technologies Corporation and Becton, Dickinson and Company through its BD Biosciences business, to use the fluorescent beads in this product only as standards, controls and calibrators for: (a) absolute counting; (b) fluorescence intensity measurements; and (c) instrument setup procedures on flow cytometry instruments, for calibrating and performance testing of flow cytometric instruments and flow cytometric assays, and for internal research and development applications. The sale of this product is expressly conditioned on the buyer not using the fluorescent beads in this product (1) in manufacturing; (2) for therapeutic, diagnostic (except when specifically labeled for such use in connection with BD products), or prophylactic purposes; (3) to resell, even if sold for calibration or research purposes. For information on purchasing a license to this product for purposes other than described above, contact Life Technologies Corporation, Cell Analysis Business Unit, Business Development, 29851 Willow Creek Road, Eugene, OR 97402. Tel: (541) 465-8300. Fax: (541) 335-0354.