

# Technical Data Sheet

## BUV661 Mouse Anti-Human CD146

### Product Information

Material Number:	750649
Size:	50 µg
Clone:	P1H12
Alternative Name:	MCAM; MELCAM; MUC18; Gicerin; Melanoma cell adhesion molecule
Reactivity:	Human (Tested in Development)
Isotype:	Mouse IgG1, κ
Immunogen:	Human Umbilical Vein Cells
Application:	Flow cytometry (Qualified)
Concentration:	0.2 mg/ml
Workshop No.:	VIII
Entrez Gene ID:	4162
Storage Buffer:	Aqueous buffered solution containing ≤0.09% sodium azide.
Regulatory Status:	RUO

### Description

The P1H12 monoclonal antibody specifically binds to CD146. CD146 is a 118 kDa transmembrane glycoprotein also known as MCAM, MUC18, or Mel-CAM. CD146 is a member of the immunoglobulin superfamily strongly expressed by blood vessel endothelium and smooth muscle. CD146 is also expressed by angioblasts, mesenchymal stem cells, melanoma cells, intermediate trophoblasts and a subpopulation of activated T cells. The P1H12 monoclonal antibody has been reported to block endothelial cell adhesion that is observed very early in embryogenesis. It can be useful in the study of embryologic vasculogenesis. This antibody is suitable for immunohistochemical staining of acetone-fixed frozen tissue sections, immunoprecipitation and ELISA.

The antibody was conjugated to BD Horizon™ BUV661 which is part of the BD Horizon Brilliant™ Ultraviolet family of dyes. This dye is a tandem fluorochrome of BD Horizon BUV395 with an Ex Max of 348-nm and an acceptor dye with an Em Max at 661-nm. BD Horizon Brilliant BUV661 can be excited by the ultraviolet laser (355 nm) and detected with a 670/25 filter and a 630 nm LP. Due to cross laser excitation of this dye, there may be significant spillover into channels detecting APC-like emissions (eg, 670/25-nm filter).

Due to spectral differences between labeled cells and beads, using BD™ CompBeads can result in incorrect spillover values when used with BD Horizon BUV661 reagents. Therefore, the use of BD CompBeads or BD CompBeads Plus to determine spillover values for these reagents is not recommended. Different BUV661 reagents (eg, CD4 vs. CD45) can have slightly different fluorescence spillover therefore, it may also be necessary to use clone-specific compensation controls when using these reagents.

### Preparation and Storage

Store undiluted at 4°C and protected from prolonged exposure to light. Do not freeze. The monoclonal antibody was purified from tissue culture supernatant or ascites by affinity chromatography. The antibody was conjugated with BD Horizon BUV661 under optimal conditions that minimize unconjugated dye and antibody.

### Recommended Assay Procedure

For optimal and reproducible results, BD Horizon Brilliant Stain Buffer should be used anytime two or more BD Horizon Brilliant dyes (including BD OptiBuild Brilliant reagents) are used in the same experiment. Fluorescent dye interactions may cause staining artifacts which may affect data interpretation. The BD Horizon Brilliant Stain Buffer was designed to minimize these interactions. More information can be found in the Technical Data Sheet of the BD Horizon Brilliant Stain Buffer (Cat. No. 563794).

### Suggested Companion Products

Catalog Number	Name	Size	Clone
554656	Stain Buffer (FBS)	500 mL	
554657	Stain Buffer (BSA)	500 mL	

563794	Brilliant Stain Buffer	100 Tests	
555899	Lysing Buffer	100 mL	
612966	BUV661 Mouse IgG1, $\kappa$ Isotype Control	50 $\mu$ g	X40
349202	Lysing Solution 10X Concentrate	100 NA	
564219	Human BD Fc Block™	50 mg	

## Product Notices

1. This antibody was developed for use in flow cytometry.
2. The production process underwent stringent testing and validation to assure that it generates a high-quality conjugate with consistent performance and specific binding activity. However, verification testing has not been performed on all conjugate lots.
3. Researchers should determine the optimal concentration of this reagent for their individual applications.
4. An isotype control should be used at the same concentration as the antibody of interest.
5. Caution: Sodium azide yields highly toxic hydrazoic acid under acidic conditions. Dilute azide compounds in running water before discarding to avoid accumulation of potentially explosive deposits in plumbing.
6. For fluorochrome spectra and suitable instrument settings, please refer to our Multicolor Flow Cytometry web page at [wwwbdbiosciences.com/colors](http://wwwbdbiosciences.com/colors).
7. Please refer to [wwwbdbiosciences.com/us/s/resources](http://wwwbdbiosciences.com/us/s/resources) for technical protocols.
8. BD Horizon Brilliant Stain Buffer is covered by one or more of the following US patents: 8,110,673; 8,158,444; 8,575,303; 8,354,239.
9. BD Horizon Brilliant Ultraviolet 661 is covered by one or more of the following US patents: 8,110,673; 8,158,444; 8,227,187; 8,575,303; 8,354,239.

## References

- Elshal MF, Khan SS, Takahashi Y, Solomon MA, McCoy JP, Jr. CD146 (Mel-CAM), an adhesion marker of endothelial cells, is a novel marker of lymphocyte subset activation in normal peripheral blood. *Blood*. 2005; 106(8):2923-2924. (Clone-specific: Flow cytometry).
- Kishimoto T, Tadimitsu Kishimoto .. et al., ed. Leucocyte typing VI : white cell differentiation antigens : proceedings of the sixth international workshop and conference held in Kobe, Japan, 10-14 November 1996. New York: Garland Pub.; 1997; .
- Sers C, Kirsch K, Rothbacher U, Riethmüller G, Johnson JP. Genomic organization of the melanoma-associated glycoprotein MUC18: implications for the evolution of the immunoglobulin domains. *Proc Natl Acad Sci U S A*. 1993; 90(18):8514-8518. (Biology: Flow cytometry).
- Shih IM, Elder DE, Hsu MY, Herlyn M. Regulation of Mel-CAM/MUC18 expression on melanocytes of different stages of tumor progression by normal keratinocytes. *Am J Pathol*. 1994; 145(4):837-845. (Biology: Flow cytometry).
- Shih IM. The role of CD146 (Mel-CAM) in biology and pathology. *J Pathol*. 1999; 189(1):4-11. (Biology: Flow cytometry).
- Solovey A, Lin Y, Browne P, Choong S, Wayner E, Hebbel R P. Circulating activated endothelial cells in sickle cell anemia. *N Engl J Med*. 1997; 337(22):1584-1590. (Immunogen: Flow cytometry).
- Solovey AN, Gui L, Chang L, Enestein J, Browne PV, Hebbel RP. Identification and functional assessment of endothelial P1H12. *J Lab Clin Med*. 2001; 138(5):322-331. (Clone-specific: Flow cytometry).

## BD Biosciences

[bdbiosciences.com](http://bdbiosciences.com)

United States 877.232.8995	Canada 888.268.5430	Europe 32.53.720.550	Japan 0120.8555.90	Asia Pacific 65.6861.0633	Latin America/Caribbean 0800.771.7157
-------------------------------	------------------------	-------------------------	-----------------------	------------------------------	--

For country contact information, visit [bdbiosciences.com/contact](http://bdbiosciences.com/contact)

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for a patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton Dickinson and Company is strictly prohibited. For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

©2020 BD. All rights reserved. Unless otherwise noted, BD, the BD Logo and all other trademarks are the property of Becton, Dickinson and Company or its affiliates.

